

# ChM<sup>®</sup>
































5.0 ChM Locked Plating  
ChLP system

5.0ChLP scapula plates  
3.7207; 3.7208  
3.7209; 3.7210  
3.7211; 3.7212

- *SURGICAL TECHNIQUE*
- *IMPLANTS*
- *INSTRUMENT SET*



## SYMBOLS DESCRIPTION

	Titanium or titanium alloy		H length [mm]
	Cobalt		Angle
	Left		available lengths
	Right		Available number of holes
	Available versions: left/right		Thickness [mm]
	Length		Scale 1:1
	Torx drive		Number of threaded holes in the shaft part of the plate
	Torx drive cannulated		Number of locking holes in the plate
	Hexagonal drive		Variable angle
	Hexagonal drive cannulated		Cortical
	Cannulated		Cancellous
	Locking		Available in sterile/ non- sterile condition
	Diameter [mm]		Refer to surgical technique
	Caution - pay attention to a special procedure.		
	Perform the activity under X-Ray control.		
	Information about the next stages of a procedure.		
	Proceed to the next stage.		
	Return to the specified stage and repeat the activity.		
	Before using the product, carefully read the Instructions for Use. It contains, among others, indications, contraindications, side effects, recommendations and warnings related to the use of the product.		
	The above description is not a detailed instruction of conduct. The surgeon decides about choosing the operating procedure.		

**www.chm.eu**

Document No ST/80-519  
 Date of issue 21.09.2018  
 Review date P-001-10.12.2020

*The manufacturer reserves the right to introduce design changes.*

*Updated INSTRUCTIONS FOR USE are available at the following website: ifu.chm.eu*

<b>1. INTRODUCTION</b>	<b>5</b>
<b>2. IMPLANT FEATURES</b>	<b>6</b>
<b>3. SURGICAL TECHNIQUE</b>	<b>8</b>
3.1. PATIENT'S POSITIONING	8
3.2. SURGICAL APPROACH	8
3.3. FRACTURE REDUCTION	9
3.4. IMPLANT SELECTION	9
3.5. PLATE INSERTION	9
3.6. TEMPORARY PLATE STABILIZATION	9
3.7. CORTICAL SCREWS INSERTION	10
3.8. LOCKING SCREWS INSERTION	10
3.9. WOUND CLOSURE	10
<b>4. SURGICAL PROCEDURES</b>	<b>11</b>
4a. PROCEDURE OF TEMPORARY IMPLANT STABILIZATION	11
4b. PROCEDURE OF CORTICAL SELF-TAPPING SCREW 3.5 [3.1306] INSERTION	12
4c. PROCEDURE OF 5.0ChLP SELF-TAPPING SCREW 3.5 [3.5200] INSERTION	13
4d. PROCEDURE OF 5.0ChLP SCREW VA 3.5 [4.5236] INSERTION	14
<b>5. POSTOPERATIVE PROCEDURE</b>	<b>16</b>
<b>6. IMPLANT REMOVAL</b>	<b>16</b>
<b>7. CATALOGUE PAGES</b>	<b>17</b>
7a. INSTRUMENT SET	17
7b. IMPLANTS	19
7c. SCREWS	22





## 1. INTRODUCTION

This surgical technique applies to 5.0ChLP locked plating system used for stabilization of scapula fractures. The plates are a part of the ChLP locked plating system developed by **ChM**. The presented range of implants is made of materials in accordance with ISO 5832 standards. Compliance with the requirements of quality management systems and the requirements of Directive 93/42/EEC concerning medical devices guarantee high quality of the offered implants.

The system includes:

- implants (*plates and screws*),
- instrument set used in the surgery,
- surgical technique.

### Indications

---

- Complex glenoid and scapula border fractures,
- Mal-unions and non-unions.

### Plate selection and shaping

---

The plates are available in different lengths, separately for right and left side. This allows for optimal selection of the implant to the fracture type. Shaping of the plates in their epiphyseal part is not allowed.



Before using the product, carefully read the Instructions for Use. It contains, among others, indications, contraindications, side effects, recommendations and warnings related to the use of the product.




The above description is not a detailed instruction of conduct. The surgeon decides about choosing the operating procedure.

---

## 2. IMPLANT FEATURES

Scapula plates are a part of 5.0ChLP system. This system includes also compatible locking screws. To facilitate the identification, both titanium plate and screws are brown anodized.

### 5.0ChLP lateral scapula plate




### 5.0ChLP glenoid plate



### 5.0ChLP medial scapula plate






### 3. SURGICAL TECHNIQUE

#### 3.1. PATIENT'S POSITIONING

The patient may be positioned in either the lateral decubitus position or prone. Visualization of the affected limb under fluoroscopy must be ensured.



decubitus position




prone position

#### 3.2. SURGICAL APPROACH

##### 5.0ChLP MEDIAL SCAPULA PLATE


##### 5.0ChLP LATERAL SCAPULA PLATE

Posterior (*Judet*) approach. The skin incision is performed along the superior border of the scapula spine then curved to the medial scapula border and ending at the inferior angle of the scapula.



##### 5.0ChLP GLENOID PLATE

Posterior approach. Vertical, about 10cm long skin incision is performed downwards, starting at the tip of the acromion posteriorly. Particular attention should be paid to the suprascapular nerve which has to be exposed.





### 3.3. FRACTURE REDUCTION

Perform fracture reduction. If need be, temporarily stabilize the bone fragments with Kirschner wires and/or reduction pliers.

### 3.4. IMPLANT SELECTION

Select the right size of an implant to the type of fracture, bone size and structure.



### 3.5. PLATE INSERTION

Position the implant correctly on the bone.




### 3.6. TEMPORARY PLATE STABILIZATION

Stabilize the position of the implant inserting Kirschner wires into appropriate holes or using setting-compressing screw (acc. to procedure 4a).



### 3.7. CORTICAL SCREWS INSERTION

Insert cortical self-tapping screw 3.5 **[3.1306]** into the oval-shaped hole of the plate.



### 3.8. LOCKING SCREWS INSERTION

Insert locking screws, of a suitable length, into the locking holes of the plate.

- 5.0ChLP self-tapping screw 3.5 **[3.5200]** (acc. to procedure 4c).
- 5.0ChLP screw VA 3.5 **[3.5236]** (acc. to procedure 4d).



Insert the cortical screws 3.5 into the fracture before inserting the locking screws.



The doctor decides about the order and number of locking and cortical screws to be inserted.

### 3.9. WOUND CLOSURE

Before closing the wound, take an X-Ray image in at least two projections to confirm implant position and fracture reduction.

Make sure all the screws are properly tightened and do not penetrate the joint surface.


Use appropriate surgical technique to close the wound.

## 4. SURGICAL PROCEDURES

### 4a. PROCEDURE OF TEMPORARY IMPLANT STABILIZATION



#### Stabilization using Kirschner wires

- Stabilize temporary the implant inserting Kirschner wires 1.5/210 **[40.4592.210]** into dedicated holes in the plate.





#### Stabilization in locking holes using Kirschner wires

- Insert guide sleeve 5.0/1.8 **[40.5673.718]** into the locking hole of the plate.
- Insert Kirschner wire **[40.4592.210]** through the guide sleeve 5.0/1.8 **[40.5673.718]**.



#### Stabilization using setting-compressing screw

- Insert guide sleeve 5.0/2.8 **[40.5673.728]** into the locking hole of the plate.
- Insert setting-compressing screw 2.8/180 **[40.5674.728]** through the guide sleeve 5.0/2.8 **[40.5673.728]**.
- Tighten the nut of the setting-compressing screw **[40.5674.728]** and push the plate to the bone.



## 4b. PROCEDURE OF CORTICAL SELF-TAPPING SCREW 3.5 [3.1306] INSERTION

### Compression guide positioning

Position the compression guide 2.5 [40.4804.725] in a desired position:



**NEUTRAL POSITION:** Push the guide to the plate. It will position itself so as neutral insertion of the screw is allowed.

**COMPRESSION POSITION:** Do not push the guide and move it to the edge of the compression hole. The hole drilled in this position allows compressive insertion of the screw.

**ANGULAR POSITION:** Angular position of the guide may also be applied.

NEUTRAL POSITION



COMPRESSION POSITION

ANGULAR POSITION

### Hole drilling

Perform a hole through both cortices for a cortical screw 3.5 insertion. For drilling, use drill with scale 2.5/210 [40.5912.212] and compression guide in a desired position.






### Measurement of hole depth

Insert depth measure [40.4639.550] into drilled hole until the hook of the measure rests against the outer surface of the second cortex.



### Screw insertion

Insert cortical screw using handle ratchet device [40.6654.000] and screwdriver tip T15 [40.5677.000].





## 4c. PROCEDURE OF 5.0ChLP SELF-TAPPING SCREW 3.5 [3.5200] INSERTION

### Guide sleeve insertion

- Insert guide sleeve 5.0/2.8 [40.5673.728] into a locking hole of the plate.



40.5673.728

### Hole drilling

Drill using drill with scale 2.8/210 [40.5653.212] until desired depth is reached.



40.5653.212

### Measurement of hole depth

**OPTION I:** Read the length of the screw from the drill measure [40.5653.212]



40.5653.212

**OPTION II:** or use screw length measure [40.5675.500].



40.5675.500

**OPTION III:** Having removed the guide sleeve 5.0/2.8 [40.5673.728], use depth measure [40.4639.550] to determine the length of a screw.



40.4639.550

### Screw insertion

Remove the guide sleeve 5.0/2.8 [40.5673.728]. Use torque limiting ratchet handle 2Nm [40.6652.000] and screwdriver tip T15 [40.5677.000] to insert the locking screw.




40.6652.000



40.5677.000



The final tightening of the locking screw, especially when a drive is used, should always be performed with the use of torque limiting handle. Failure to use the torque limiting handle may lead to intraoperative and postoperative complications (*during later removal of the plate and locking screws*).



## 4d. PROCEDURE OF 5.0ChLP SCREW VA 3.5 [4.5236] INSERTION



When using variable angle (VA) screws, there is a risk of collision of screws or a drill with already implanted screws. Well-thought-out trajectory of inserted screws and intraoperative X-Ray control of drilling reduces the risk of the collision.

## Guide VA positioning

- Insert the guide VA 2.8 [40.8206.028] into the locking hole co-axially.
- Set the desired inclination of the guide in relation to the locking hole axis. The guide enables the inclination of 15° in each direction with respect to the axis of the locking hole.



40.8206.028



Exceeding the inclination angle of more than 15° may prevent proper locking of the VA screw in the plate hole.

## Hole drilling

- Drill using drill with scale 2.8/210 [40.5653.212] until desired depth is reached.



40.5653.212



Drill under X-Ray control to avoid a collision of the drill with already implanted screws.

## Measurement of hole depth

OPTION I: Read the length of the screw from the drill measure [40.5653.212].



40.5653.212

OPTION II: or use screw length measure [40.5675.500].




40.5675.500

OPTION III: Having removed the guide VA, use depth measure [40.4639.550] to determine the length of the screw.



40.4639.550



## Screw insertion


Use torque limiting ratchet handle 2Nm **[40.6652.000]** and screwdriver tip T15 **[40.5677.000]** to insert the VA screw.



40.6652.000



40.5677.000




When using torque limiting handle to tighten the VA screw with large inclination in relation to the axis of the locking hole, the head of the screw may protrude above the plate. In this case, it may be necessary to use a handle ratchet device **[40.6654]** and screwdriver tip T15 **[40.5677]**. Use the instruments carefully to tighten the VA screw. Avoid damaging the screw socket or screwdriver tip. Do not insert the screw too deep into the plate.



40.6654.000



40.5677.000



### Change of the VA screw positioning



It is possible to lock the VA screw three times in the threaded hole of the plate. The hole in the plate in which the VA screw was locked cannot be used to insert a standard locking screw.

## 5. POSTOPERATIVE PROCEDURE

Introduce appropriate post-operative treatment. The physician decides on the post-operative treatment and its conduct. In order to avoid patient's movement limitations, introduce exercises as soon after surgery as possible. However, make sure that the limb is not fully loaded before fragments osteosynthesis is complete.

## 6. IMPLANT REMOVAL















The physician decides about implant removal. In order to remove the implants from the body, unlock all the locking screws first and then remove them from the bone. This will prevent any rotation of the plate when removing the last locking screw.









## 7. CATALOGUE PAGES

## 7a. INSTRUMENT SET

Instrument set for 5.0ChLP 4x4 1/2H

15.0205.206

	Name	Catalogue No.	Pcs
	Tray for 5.0ChLP instrument set 4x4 1/2H	<b>14.0205.206</b>	1
	Kirschner wire 1.5/210	<b>40.4592.210</b>	4
	Drill 1.8/210	<b>40.2063.212</b>	2
	Drill with scale 2.5/210	<b>40.5912.212</b>	2
	Drill with scale 2.8/210	<b>40.5653.212</b>	2
	Screwdriver tip T15	<b>40.5677.000</b>	1
	Torque limiting ratchet handle 2Nm	<b>40.6652.000</b>	1
	Handle ratchet device	<b>40.6654.000</b>	1
	Protective guide 7/5	<b>40.5672.000</b>	2
	Guide VA 2.8	<b>40.8206.028</b>	1
	Compression guide 2.5	<b>40.4804.725</b>	1
	Guide sleeve 5.0/1.8	<b>40.5673.718</b>	2
	Guide sleeve 5.0/2.8	<b>40.5673.728</b>	4
	Depth measure	<b>40.4639.550</b>	1

	Name	Catalogue No.	Pcs
	Tray for 5.0ChLP instrument set 4x4 1/2H	<b>14.0205.202</b>	1
	Setting-compressing screw 2.8/180	<b>40.5674.728</b>	1
	Screw length measure	<b>40.5675.500</b>	1
	Plates bender 5.0	<b>40.4643.500</b>	2
	Tripod screwdriver tip 5.0ChLP	<b>40.6271.500</b>	1
	T15 screwdriver tip with holder	<b>40.6254.000</b>	1
	Cortical tap HA 3.5 with handle	<b>40.2548.200</b>	1
	Tap 5.0ChLP-3.5	<b>40.5661.000</b>	1
<b>Optional instrument</b>			
	Torque connector 2Nm	<b>40.5927.020</b>	1





7b. IMPLANTS



5.0ChLP medial scapula plate

	Len	L	R
3	72	3.7207.603	3.7208.603
5	91	3.7207.605	3.7208.605

O - holes number in shaft part of the plate



Tray for plates 5.0ChLP 3.7207-3.7212 4x4 1/2H



14.0205.426



5.0ChLP lateral scapula plate

	Len	L	R
5	91	3.7209.605	3.7210.605
7	111	3.7209.607	3.7210.607

O - holes number in shaft part of the plate







5.0ChLP glenoid plate







91




3.7211.600



3.7212.600



## 5.0ChLP self-tapping screw 3.5



Len	Ster Non Ster
12	3.5200.012
14	3.5200.014
16	3.5200.016
18	3.5200.018
20	3.5200.020
22	3.5200.022
24	3.5200.024
26	3.5200.026
28	3.5200.028
30	3.5200.030
32	3.5200.032
34	3.5200.034
36	3.5200.036
38	3.5200.038
40	3.5200.040
42	3.5200.042
44	3.5200.044
46	3.5200.046
48	3.5200.048
50	3.5200.050
52	3.5200.052
54	3.5200.054
56	3.5200.056
58	3.5200.058
60	3.5200.060
65	3.5200.065
70	3.5200.070
75	3.5200.075
80	3.5200.080
85	3.5200.085

## 5.0ChLP screw VA 3.5



Len	Co
12	4.5236.012
14	4.5236.014
16	4.5236.016
18	4.5236.018
20	4.5236.020
22	4.5236.022
24	4.5236.024
26	4.5236.026
28	4.5236.028
30	4.5236.030
32	4.5236.032
34	4.5236.034
36	4.5236.036
38	4.5236.038
40	4.5236.040
42	4.5236.042
44	4.5236.044
46	4.5236.046
48	4.5236.048
50	4.5236.050
52	4.5236.052
54	4.5236.054
56	4.5236.056
58	4.5236.058
60	4.5236.060
65	4.5236.065
70	4.5236.070
75	4.5236.075
80	4.5236.080
85	4.5236.085
90	4.5236.090
95	4.5236.095

## Cortical self-tapping screw 3.5



Len	Ster Non Ster
10	3.1306.010
12	3.1306.012
14	3.1306.014
16	3.1306.016
18	3.1306.018
20	3.1306.020
22	3.1306.022
24	3.1306.024
26	3.1306.026
28	3.1306.028
30	3.1306.030
32	3.1306.032
34	3.1306.034
36	3.1306.036
38	3.1306.038
40	3.1306.040
45	3.1306.045
50	3.1306.050
55	3.1306.055
60	3.1306.060
65	3.1306.065
70	3.1306.070
75	3.1306.075
80	3.1306.080
85	3.1306.085



**ChM sp. z o.o.**

Lewickie 3b  
16-061 Juchnowiec Kościelny  
Polska  
tel. +48 85 86 86 100  
fax +48 85 86 86 101  
chm@chm.eu  
www.chm.eu



CE 0197